

Supplementary Table 3. Differences in the atrial and ventricular impedance, lead sensing, and captured threshold values between the MR-conditional and conventional groups at pre-/post-MRI and follow-up

	MR-conditional (n = 16)	Conventional (n = 27)	p value
Impedance, Ω			
Atrial			
Pre-MRI	466.4 ± 49.7	444.1 ± 66.3	0.35
Post-MRI	463.8 ± 44.4	449.3 ± 61.4	0.52
Follow-up	458.5 ± 45.3	465.7 ± 58.9	0.77
Ventricular			
Pre-MRI	539.1 ± 73.8	587.4 ± 139.2	0.22
Post-MRI	542.4 ± 72.9	586.1 ± 135.9	0.24
Follow-up	543.6 ± 79.1	556.5 ± 102.3	0.72
Sensing, mV			
Atrial			
Pre-MRI	2.24 ± 1.18	2.48 ± 1.09	0.67
Post-MRI	2.19 ± 1.12	2.87 ± 1.59	0.31
Follow-up	1.78 ± 0.94	2.84 ± 1.93	0.15
Ventricular			
Pre-MRI	11.1 ± 3.88	12.6 ± 5.86	0.39
Post-MRI	11.4 ± 3.89	12.9 ± 5.82	0.36
Follow-up	12.5 ± 4.47	10.7 ± 5.65	0.33
Capture threshold, V			
Atrial			
Pre-MRI	0.53 ± 0.15	0.71 ± 0.29	0.08
Post-MRI	0.53 ± 0.13	0.54 ± 0.18	0.89
Follow-up	0.5 ± 0	0.74 ± 0.36	0.05
Ventricular			
Pre-MRI	0.71 ± 0.23	0.68 ± 0.27	0.67
Post-MRI	0.73 ± 0.16	0.73 ± 0.26	0.96
Follow-up	0.76 ± 0.24	0.71 ± 0.24	0.53

Values are presented as mean ± SD.

MR, magnetic resonance; MRI, magnetic resonance imaging.